

Certificate of Analysis



Lot status: APPROVED

MODEL	Petri ø90mm Single Wrapped	PRODUCT CODE	MU024/10
MEDIA	CETRIMIDE	MEDIA COMPOSITION	Cetrimide Agar
BATCH	82403	STORAGE CONDITIONS	at 2-25°C until expiry date
DATE OF MANUFACTURING	20 Feb 2024	EXPIRY DATE	20 Aug 2024

Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Light yellow coloured, clear to slightly opalescent gel forms in plates	Conforms
pH specification	7.2±0.2	Conforms
Filling media volume	30ml±2ml	Conforms
Packaging	Single Wrapped (S.W.)	Conforms

MICROBIOLOGICAL TEST CONTROLS

- Sterility control**

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 120 hours	No growth	Conforms
22.5±2.5°C aerobic for 120 hours		

- Fertility – Growth Promotion Test** according to USP/EP

Control strain	Medium inoculation level	Incubation Conditions	Recovery Specifications	Recovery Results	Cultural Response
<i>P. aeruginosa</i> ATCC 9027	10-100 viable microorganisms	18 h at 32.5 ± 2.5°C	Good growth. Colonies may appear pigmented yellow-green or nonpigmented	Good growth. Colonies may appear pigmented yellow-green or nonpigmented	Conforms
<i>E. coli</i> ATCC 8739	≥100 viable microorganisms	72 h at 32.5 ± 2.5°C	Inhibited	Inhibited	Conforms

Start analysis	20 Feb 2024	End analysis	26 Feb 2024
Quality Control	Elisabetta Peri	<i>Elisabetta Peri</i>	
Quality Assurance	Celeste Annovazzi	<i>Celeste Annovazzi</i>	