

# Certificate of Analysis



Lot status: **APPROVED**

**MODEL:** 1000ml Screw Cap Bottle  
**MEDIA:** Peptone Water 1 + Tween80 0.5% pH 7.2  
**BATCH:** 23B0602  
**DATE OF MANUFACTURING:** 06 Feb 2023

**PRODUCT CODE:** 001.1/33PSC28.1L  
**STORAGE CONDITIONS:** at 2-25°C until expiry date  
**EXPIRY DATE:** 06 Feb 2024

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Colourless clear solution. White crystals may appear. Swirl to dissolve.	Conforms
pH specification	7.2±0.2	Conforms
Filling media volume	1000ml	Conforms
Media composition	Casein Peptone 1.00g/L + NaCl 8.50g/L + Tween80 0.5%	Conforms
	Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer	

## MICROBIOLOGICAL TEST CONTROLS

- **Sterility control** according to USP/EP

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 336 hours 22.5±2.5°C aerobic for 336 hours	No growth	Conforms

Start analysis	07 Feb 2023	End analysis	21 Feb 2023
Quality Control	Claudia Previtali	<i>Claudia Previtali</i>	
Quality Assurance	Rosaria Cassese	<i>Rosaria Cassese</i>	

**GHIARONI CODE: 9538281000**