

# Certificate of Analysis



Lot status: **APPROVED**

**MODEL:** 900ml Bottle with Pierceable Cap  
**MEDIA:** Peptone Water 1 + Tween80 0.1% pH 7.0

**PRODUCT CODE:** 001.2/33P.900

**BATCH:** 23E1702

**STORAGE CONDITIONS:** at 2-25°C until expiry date

**DATE OF MANUFACTURING:** 17 May 2023

**EXPIRY DATE:** 17 May 2024

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Colourless clear solution.	Conforms
pH specification	7.0±0.2	Conforms
Filling media volume	900ml	Conforms
Media composition	Casein peptone 1.00g/L + Sodium chloride 8.50g/l + Tween80 0.1%	Conforms
	Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer	

## MICROBIOLOGICAL TEST CONTROLS

- Sterility control according to USP/EP

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 336 hours 22.5±2.5°C aerobic for 336 hours	No growth	Conforms

Start analysis	19 May 2023	End analysis	05 Jun 2023
Quality Control	Claudia Previtali	<i>Claudia Previtali</i>	
Quality Assurance	Rosaria Cassese	<i>Rosaria Cassese</i>	

**GHIARONI CODE: 953808024**