

Certificate of Analysis



Lot status: **APPROVED**

MODEL: 900ml Bottle with Pierceable Cap
MEDIA: Peptone Water 1 + Tween80 0.1% pH 7.0
BATCH: 23E1802
DATE OF MANUFACTURING: 18 May 2023

PRODUCT CODE: 001.2/33P.900
STORAGE CONDITIONS: at 2-25°C until expiry date
EXPIRY DATE: 18 May 2024

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Colourless clear solution.	Conforms
pH specification	7.0±0.2	Conforms
Filling media volume	900ml	Conforms
Media composition	Casein peptone 1.00g/L + Sodium chloride 8.50g/l + Tween80 0.1% Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer	Conforms

MICROBIOLOGICAL TEST CONTROLS

- Sterility control according to USP/EP

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 336 hours 22.5±2.5°C aerobic for 336 hours	No growth	Conforms

Start analysis	19 May 2023	End analysis	05 Jun 2023
Quality Control	Claudia Previtali	<i>Claudia Previtali</i>	
Quality Assurance	Rosaria Cassese	<i>Rosaria Cassese</i>	

GHIARONI CODE: 953808024