

Certificate of Analysis



Lot status: **APPROVED**

MODEL	900ml Bottle with Pierceable Cap	PRODUCT CODE	001.2/33P.900
MEDIA	Peptone Water 1 + Tween80 0.1% pH 7.0	MEDIA COMPOSITION	Casein peptone 1.00g/L + Sodium Chloride 8.50g/l + Tween80 0.1%
BATCH	24C0101	STORAGE CONDITIONS	at 2-25°C until expiry date
DATE OF MANUFACTURING:	01 Mar 2024	EXPIRY DATE	01 Mar 2025

Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Colourless clear solution.	Conforms
pH specification	7.0±0.2	Conforms
Filling media volume	900ml	Conforms

MICROBIOLOGICAL TEST CONTROLS

- Sterility control** according to USP/EP

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 336 hours	No growth	Conforms
22.5±2.5°C aerobic for 336 hours		

Start analysis	04 Mar 2024	End analysis	18 Mar 2024
Quality Control	Elisabetta Peri		
Quality Assurance	Alessia Diego		

GHIARONI CODE: 953808024