

# Certificate of Analysis



Lot status: **APPROVED**

<b>MODEL</b>	<b>900ml Bottle with Pierceable Cap</b>	<b>PRODUCT CODE</b>	<b>001.2/33P.900</b>
<b>MEDIA</b>	Peptone Water 1 + Tween80 0.1% pH 7.0	<b>MEDIA COMPOSITION</b>	Casein peptone 1.00g/L + Sodium Chloride 8.50g/l + Tween80 0.1%
<b>BATCH</b>	<b>24C0401</b>	<b>STORAGE CONDITIONS</b>	at 2-25°C until expiry date
<b>DATE OF MANUFACTURING:</b>	<b>04 Mar 2024</b>	<b>EXPIRY DATE</b>	<b>04 Mar 2025</b>

Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Colourless clear solution.	<i>Conforms</i>
pH specification	7.0±0.2	<i>Conforms</i>
Filling media volume	900ml	<i>Conforms</i>

## MICROBIOLOGICAL TEST CONTROLS

- Sterility control** according to USP/EP

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 336 hours	No growth	<i>Conforms</i>
22.5±2.5°C aerobic for 336 hours		

Start analysis	05 Mar 2024	End analysis	19 Mar 2024
Quality Control	Elisabetta Peri	<i>Elisabetta Peri</i>	
Quality Assurance	Alessia Diego	<i>Alessia Diego</i>	

**GHIARONI CODE: 953808024**