

Certificate of Analysis



Lot status: **APPROVED**

MODEL	100ml Screw Cap Bottle	PRODUCT CODE	002.1/30PSC28.100
MEDIA	Phosphate Buffer 1 (0.25M) pH 7.2	MEDIA COMPOSITION	Monobasic Potassium Phosphate 34.00g/L
BATCH	23L1402	STORAGE CONDITIONS	at 2-25°C until expiry date
DATE OF MANUFACTURING	14 Dec 2023	EXPIRY DATE	14 Dec 2024

Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Colourless clear solution. White crystals may appear. Swirl to dissolve.	<i>Conforms</i>
pH specification	7.2±0.1	<i>Conforms</i>
Filling media volume	100ml	<i>Conforms</i>

MICROBIOLOGICAL TEST CONTROLS

- **Sterility control** according to USP/EP

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 336 hours 22.5±2.5°C aerobic for 336 hours	No growth	<i>Conforms</i>

Start analysis	18 Dec 2023	End analysis	02 Jan 2024
Quality Control	Elisabetta Peri		
Quality Assurance	Alessia Diego		