

Certificate of Analysis



Lot status: **APPROVED**

MODEL	1000ml Screw Cap Bottle	PRODUCT CODE	002.5/33PSC28.1L
MEDIA	Phosphate Buffer 5 (0.05M) pH 8.0	MEDIA COMPOSITION	Monobasic Potassium Phosphate 6.80 g/L
BATCH	23J1601	STORAGE CONDITIONS	at 2-25°C until expiry date
DATE OF MANUFACTURING	16 Oct 2023	EXPIRY DATE	16 Oct 2024

Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Colourless, clear to slightly hazy, with or without precipitate.	Conforms
pH specification	8.0±0.1	Conforms
Filling media volume	1000ml	Conforms

MICROBIOLOGICAL TEST CONTROLS

- **Sterility control** according to USP/EP

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 336 hours 22.5±2.5°C aerobic for 336 hours	No growth	Conforms

Start analysis	17 Oct 2023	End analysis	31 Oct 2023
Quality Control	Elisabetta Peri		
Quality Assurance	Alessia Diego		

GHIARONI CODE: 953821548