

# Certificate of Analysis



Lot status: **APPROVED**

<b>MODEL</b>	<b>1000ml Screw Cap Bottle</b>	<b>PRODUCT CODE</b>	<b>002.5/33PSC28.1L</b>
<b>MEDIA</b>	Phosphate Buffer 5 (0.05M) pH 8.0	<b>MEDIA COMPOSITION</b>	Monobasic Potassium Phosphate 6.80 g/L
<b>BATCH</b>	<b>24C2503</b>	<b>STORAGE CONDITIONS</b>	at 2-25°C until expiry date
<b>DATE OF MANUFACTURING</b>	<b>25 Mar 2024</b>	<b>EXPIRY DATE</b>	<b>25 Mar 2025</b>

Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Colourless, clear to slightly hazy, with or without precipitate.	Conforms
pH specification	8.0±0.1	Conforms
Filling media volume	1000ml	Conforms

## MICROBIOLOGICAL TEST CONTROLS

- **Sterility control** according to USP/EP

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 336 hours 22.5±2.5°C aerobic for 336 hours	No growth	Conforms

<b>Start analysis</b>	<b>26 Mar 2024</b>	<b>End analysis</b>	<b>09 Apr 2024</b>
<b>Quality Control</b>	Elisabetta Peri		
<b>Quality Assurance</b>	Alessia Diego		

**GHIARONI CODE: 953821548**