

Certificate of Analysis



Lot status: **APPROVED**

MODEL	200ml Screw Cap Bottle	PRODUCT CODE	004.4/31PSC28.200
MEDIA	Medium A (EP) pH 8.0	MEDIA COMPOSITION	Peptone 6.00g/L + Yeast extract 3.00g/L + Dextrose 1.00g/L + Casein peptone 4.00g/L + Beef extract 1.50g/L + Agar 15.00g/L
BATCH	24A2405	STORAGE CONDITIONS	at 2-25°C until expiry date
DATE OF MANUFACTURING	24 Jan 2024	EXPIRY DATE	24 Jan 2025

Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Amber coloured, clear to slightly opalescent gel forms in plates	Conforms
pH specification	8.0±0.1	Conforms
Filling media volume	200ml	Conforms

MICROBIOLOGICAL TEST CONTROLS

- Sterility control**

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 120 hours 22.5±2.5°C aerobic for 120 hours	No growth	Conforms

Start analysis	24 Jan 2024	End analysis	29 Jan 2024
Quality Control	Elisabetta Peri		
Quality Assurance	Alessia Diego		

GHIARONI CODE: 953821516