

Certificate of Analysis



Lot status: **APPROVED**

MODEL: 200ml Screw Cap Bottle

MEDIA: Medium B (EP) pH 7.3

BATCH: 22K2506

DATE OF MANUFACTURING: 25 Nov 2022

PRODUCT CODE: 004.5/31PSC28.200

STORAGE CONDITIONS: at 2-25°C until expiry date

EXPIRY DATE: 25 Nov 2023

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Amber coloured, clear to slightly opalescent gel forms in plates	<i>Conforms</i>
pH specification	7.3±0.1	<i>Conforms</i>
Filling media volume	200ml	<i>Conforms</i>
Media composition	Casein Peptone 17.00g/L + Soy Peptone 3.00g/L+ Dextrose 2.50g/L + Sodium Chloride 5.00g/L+ Tween80 1%+ Dibasic potassium phosphate 2.50g/L +Agar 15.00g/L	<i>Conforms</i>
	Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer	

MICROBIOLOGICAL TEST CONTROLS

- Sterility control**

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 120 hours 22.5±2.5°C aerobic for 120 hours	No growth	<i>Conforms</i>

Start analysis	29 Nov 2022	End analysis	05 Dec2022
Quality Control	Claudia Previtali	<i>Claudia Previtali</i>	
Quality Assurance	Rosaria Cassese	<i>Rosaria Cassese</i>	