

Certificate of Analysis



Lot status: **APPROVED**

MODEL	200ml Screw Cap Bottle	PRODUCT CODE	004.5/31PSC28.200
MEDIA	Medium B (EP) pH 7.3	MEDIA COMPOSITION	Casein Peptone 17.00g/L + Soy Peptone 3.00g/L+ Dextrose 2.50g/L + Sodium Chloride 5.00g/L+ Tween80 1%+ Dibasic potassium phosphate 2.50g/L +Agar 15.00g/L
BATCH	24D0302	STORAGE CONDITIONS	at 2-25°C until expiry date
DATE OF MANUFACTURING	03 Apr 2024	EXPIRY DATE	03 Apr 2025

Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Amber coloured, clear to slightly opalescent gel forms in plates	Conforms
pH specification	7.3±0.1	Conforms
Filling media volume	200ml	Conforms

MICROBIOLOGICAL TEST CONTROLS

- Sterility control**

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 120 hours 22.5±2.5°C aerobic for 120 hours	No growth	Conforms

Start analysis	03 Apr 2024	End analysis	08 Apr 2024
Quality Control	Elisabetta Peri		
Quality Assurance	Alessia Diego		

GHIARONI CODE: 953821517