

Certificate of Analysis



Lot status: **APPROVED**

MODEL:	200ml Screw Cap Bottle	PRODUCT CODE:	004.6/31PSC28.200
MEDIA:	Medium F (EP) – Medium 19 (USP) pH 6.1	STORAGE CONDITIONS:	at 2-25°C until expiry date
BATCH:	23F2703	EXPIRY DATE:	27 Jun 2024
DATE OF MANUFACTURING:	27 Jun 2023		

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Amber coloured, clear to slightly opalescent gel forms in plates	<i>Conforms</i>
pH specification	6.1±0.1	<i>Conforms</i>
Filling media volume	200ml	<i>Conforms</i>
Media composition	Peptone 9.40g/L + Yeast extract 4.70g/L + Beef extract 2.40g/L + Dextrose 10.00g/L + Sodium Chloride 10.00g/L + Agar 23.50g/L	<i>Conforms</i>
	Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer	

MICROBIOLOGICAL TEST CONTROLS

- Sterility control**

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 120 hours 22.5±2.5°C aerobic for 120 hours	No growth	<i>Conforms</i>

Start analysis	27 Jun 2023	End analysis	03 Jul 2023
Quality Control	Claudia Previtali	<i>Claudia Previtali</i>	
Quality Assurance	Rosaria Cassese	<i>Rosaria Cassese</i>	

GHIARONI CODE: 953821519