

Certificate of Analysis



Lot status: **APPROVED**

MODEL: 200ml Screw Cap Bottle
MEDIA: Medium G (EP) pH 7.0

PRODUCT CODE: 004.7/31PSC28.200

BATCH: 22F1603

STORAGE CONDITIONS: at 2-25°C until expiry date

DATE OF MANUFACTURING: 16 Jun 2022

EXPIRY DATE: 16 Jun 2023

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Amber coloured, clear to slightly opalescent gel forms in plates	Conforms
pH specification	7.0±0.1	Conforms
Filling media volume	200ml	Conforms
Media composition	Peptone 10.00g/L + Meat extract 10.00g/L + Glycerol 10.00g/L + Sodium chloride 3.00g/L + Agar 15.00g/L	Conforms
	Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer	

MICROBIOLOGICAL TEST CONTROLS

- Sterility control**

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 120 hours 22.5±2.5°C aerobic for 120 hours	No growth	Conforms

Start analysis	17 Jun 2022	End analysis	22 Jun 2022
Quality Control	Claudia Previtali	<i>Claudia Previtali</i>	
Quality Assurance	Rosaria Cassese	<i>Rosaria Cassese</i>	

GHIARONI CODE: 953821520