

Certificate of Analysis



Lot status: **APPROVED**

MODEL: 200ml Screw Cap Bottle
MEDIA: Medium 10 (USP) pH 7.2

PRODUCT CODE: 004.A/31PSC28.200

BATCH: 23H0203

STORAGE CONDITIONS: at 2-25°C until expiry date

DATE OF MANUFACTURING: 02 Aug 2023

EXPIRY DATE: 02 Aug 2024

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Amber coloured, clear to slightly opalescent gel forms in plates	<i>Conforms</i>
pH specification	7.2±0.1	<i>Conforms</i>
Filling media volume	200ml	<i>Conforms</i>
Media composition	Casein peptone 17.00g/L + Soy peptone 3.00g/L + Sodium chloride 5.00g/L + Dextrose 2.50g/L + Dibasic potassium phosphate 2.50g/L + Agar 12.00g/L + Tween80 1%	<i>Conforms</i>
	Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer	

MICROBIOLOGICAL TEST CONTROLS

- Sterility control**

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 120 hours 22.5±2.5°C aerobic for 120 hours	No growth	<i>Conforms</i>

Start analysis	02 Aug 2023	End analysis	07 Aug 2023
Quality Control	Claudia Previtali	<i>Claudia Previtali</i>	
Quality Assurance	Rosaria Cassese	<i>Rosaria Cassese</i>	

GHIARONI CODE: 953821526