Certificate of Analysis

Lot status: **APPROVED**

MODEL:	**200ml Screw Cap Bottle**	**PRODUCT CODE:**	**004.C/31PSC28.200**
MEDIA:	Medium 4 (USP) pH 6.6		
BATCH:	**21D1903**	**STORAGE CONDITIONS:**	at 2-25°C until expiry date
DATE OF MANUFACTURING:	**19 Apr 2021**	**EXPIRY DATE:**	**19 Apr 2022**

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Amber coloured, clear to slightly opalescent gel forms in plates	***Conforms***
pH specification	6.6±0.1	***Conforms***
Filling media volume	200ml	***Conforms***
Media composition	Peptone 6.00g/L + Yeast extract 3.00g/L + Dextrose 1.00g/L + Beef extract 1.50g/L + Agar 15.00g/L	***Conforms***
	Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer	

MICROBIOLOGICAL TEST CONTROLS

- **Sterility control**

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 120 hours 22.5±2.5°C aerobic for 120 hours	No growth	***Conforms***

Start analisys	**21 Apr 2021**	**End analisys**	**26 Apr 2021**
Quality Control	**Stefania Nichele**	Nichele Stefania	
Quality Assurance	**Celeste Annovazzi**	Celeste Annovazzi	

GHIARONI CODE: 953821523