

Certificate of Analysis



Lot status: **APPROVED**

MODEL	1000ml Bottle with Pierceable Cap	PRODUCT CODE	006.1/33P.1L
MEDIA	NaCl Peptone Buffer pH 7.0	MEDIA COMPOSITION	Casein peptone 1.00 g/L + Sodium chloride 4.30g/L + Potassium dihydrogen phosphate 3.60g/L + Disodium hydrogen phosphate 7.20g/L
BATCH	24A2203	STORAGE CONDITIONS	at 2-25°C until expiry date
DATE OF MANUFACTURING	22 Jan 2024	EXPIRY DATE	22 Jan 2025

Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Colourless, clear to slightly hazy, with or without precipitate.	<i>Conforms</i>
pH specification	7.0±0.2	<i>Conforms</i>
Filling media volume	1000ml	<i>Conforms</i>

MICROBIOLOGICAL TEST CONTROLS

- **Sterility control** according to USP/EP

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 336 hours 22.5±2.5°C aerobic for 336 hours	No growth	<i>Conforms</i>

Start analysis	23 Jan 2024	End analysis	06 Feb 2024
Quality Control	Elisabetta Peri		
Quality Assurance	Alessia Diego		