

Certificate of Analysis



Lot status: **APPROVED**

MODEL: 900ml Bottle with Pierceable Cap
MEDIA: NaCl Peptone Buffer pH 7.0
BATCH: 23C0902
DATE OF MANUFACTURING: 09 Mar 2023

PRODUCT CODE: 006.1/33P.900
STORAGE CONDITIONS: at 2-25°C until expiry date
EXPIRY DATE: 09 Mar 2024

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Colourless, clear to slightly hazy, with or without precipitate.	Conforms
pH specification	7.0±0.2	Conforms
Filling media volume	900ml	Conforms
Media composition	Casein peptone 1.00 g/L + Sodium chloride 4.30g/L + Potassium dihydrogen phosphate 3.60g/L + Disodium hydrogen phosphate 7.20g/L Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer	Conforms

MICROBIOLOGICAL TEST CONTROLS

- **Sterility control** according to USP/EP

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 336 hours 22.5±2.5°C aerobic for 336 hours	No growth	Conforms

Start analysis	10 Mar 2023	End analysis	24 Mar 2023
Quality Control	Claudia Previtali	<i>Claudia Previtali</i>	
Quality Assurance	Rosaria Cassese	<i>Rosaria Cassese</i>	

GHIARONI CODE: 953808005