

Certificate of Analysis



Lot status: **APPROVED**

MODEL: 100ml w/Pierceable Flip-Off Cap Bottle
MEDIA: Tryptone Soy Broth
BATCH: 23B2702
DATE OF MANUFACTURING: 27 Feb 2023

PRODUCT CODE: 011/30FLP32.100
STORAGE CONDITIONS: at 2-25°C until expiry date
EXPIRY DATE: 27 Feb 2025

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	yellow coloured, clear solution	Conforms
pH specification	7.3±0.2	Conforms
Filling media volume	100ml	Conforms
Media composition	Tryptone Soy Broth	Conforms
	Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer	

MICROBIOLOGICAL TEST CONTROLS

- Sterility control**

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 336 hours 22.5±2.5°C aerobic for 336 hours	No growth	Conforms

- Fertility – Growth Promotion Test** according to USP/EP

Control strain	Medium inoculation level	Incubation Conditions	Recovery Specifications	Recovery Results	Cultural Response
<i>E. coli</i> ATCC 8739	10-100 viable microorganisms	18-72 h at 32.5 ± 2.5°C	Positive growth	Positive growth	Conforms
<i>P. aeruginosa</i> ATCC 9027	10-100 viable microorganisms	18-72 h at 32.5 ± 2.5°C	Positive growth	Positive growth	Conforms
<i>S. aureus</i> ATCC 6538	10-100 viable microorganisms	18-72 h at 32.5 ± 2.5°C	Positive growth	Positive growth	Conforms
<i>B. subtilis</i> ATCC 6633	10-100 viable microorganisms	18-72 h at 32.5 ± 2.5°C	Positive growth	Positive growth	Conforms
<i>B. subtilis</i> ATCC 6633	10-100 viable microorganisms	18-72 h at 22.5 ± 2.5°C	Positive growth	Positive growth	Conforms
<i>C. albicans</i> ATCC 10231	10-100 viable microorganisms	48-120 h at 22.5 ± 2.5°C	Positive growth	Positive growth	Conforms
<i>A. brasiliensis</i> ATCC 16404	10-100 viable microorganisms	48-120 h at 22.5 ± 2.5°C	Positive growth	Positive growth	Conforms

Start analysis	28 Feb 2023	End analysis	14 Mar 2023
Quality Control	Claudia Previtali	<i>Claudia Previtali</i>	
Quality Assurance	Rosaria Cassese	<i>Rosaria Cassese</i>	