

Certificate of Analysis



Lot status: **APPROVED**

MODEL: 1000ml Screw Cap Bottle
MEDIA: Buffered Peptone Medium (EP) pH 7.0
BATCH: 23I0702
DATE OF MANUFACTURING: 07 Sep 2023

PRODUCT CODE: 012.1/33PSC28.1L
STORAGE CONDITIONS: at 2-25°C until expiry date
EXPIRY DATE: 07 Sep 2024

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Yellow solution. Crystals may appear. Swirl to dissolve.	Conforms
pH specification	7.0±0.2	Conforms
Filling media volume	1000ml	Conforms
Media composition	Peptone 10.00g/L + Sodium Chloride 5.00g/L + Potassium dihydrogen phosphate 1.50g/L + Disodium hydrogen phosphate 3.50g/L Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer	Conforms

MICROBIOLOGICAL TEST CONTROLS

- Sterility control**

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 336 hours 22.5±2.5°C aerobic for 336 hours	No growth	Conforms

Start analysis	08 Sep 2023	End analysis	22 Sep 2023
Quality Control	Claudia Previtali	<i>Claudia Previtali</i>	
Quality Assurance	Rosaria Cassese	<i>Rosaria Cassese</i>	

GHIARONI CODE: 953820516