

Certificate of Analysis



Lot status: **APPROVED**

MODEL	1000ml Screw Cap Bottle	PRODUCT CODE	012.1/33PSC28.1L
MEDIA	Buffered Peptone Medium (EP) pH 7.0	MEDIA COMPOSITION	Peptone 10.00g/L + Sodium Chloride 5.00g/L + Potassium dihydrogen phosphate 1.50g/L + Disodium hydrogen phosphate 3.50g/L
BATCH	24C0402	STORAGE CONDITIONS	at 2-25°C until expiry date
DATE OF MANUFACTURING	04 Mar 2024	EXPIRY DATE	04 Mar 2025

Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Yellow solution. Crystals may appear. Swirl to dissolve.	Conforms
pH specification	7.0±0.2	Conforms
Filling media volume	1000ml	Conforms

MICROBIOLOGICAL TEST CONTROLS

- Sterility control**

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 336 hours 22.5±2.5°C aerobic for 336 hours	No growth	Conforms

Start analysis	05 Mar 2024	End analysis	19 Mar 2024
Quality Control	Elisabetta Peri		
Quality Assurance	Alessia Diego		

GHIARONI CODE: 953820516