

# Certificate of Analysis



Lot status: **APPROVED**

<b>MODEL</b> RODAC ø55mm Triple Wrapped Sterile Irradiated	<b>PRODUCT CODE</b> 063INHY/21
<b>MEDIA</b> SDA+Lec+T80+Hys+ThioS	<b>MEDIA COMPOSITION</b> Sabouraud Dextrose Agar + 0.07% Lecithin + 0.5% Tween80 + 0.1% L-Hystidine + 0.05% S.Thiosulphate
<b>BATCH</b> 33401	<b>STORAGE CONDITIONS</b> at 2-25°C until expiry date
<b>DATE OF MANUFACTURING</b> 07 Mar 2024	<b>EXPIRY DATE</b> 07 Sep 2024

Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	Amber coloured, clear to slightly opalescent gel forms in plates	Conforms
pH specification	5.6±0.2	Conforms
Filling media volume	17ml±1ml	Conforms
Packaging	Triple Wrapped Sterile Irradiated ( T.W.S.I.)	Conforms
Dose of irradiation	10-25 KGy	Conforms Irradiation certificate Nr: 3515345

## MICROBIOLOGICAL TEST CONTROLS

- Sterility control**

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 120 hours	No growth	Conforms
22.5±2.5°C aerobic for 120 hours		

- Fertility – Growth Promotion Test** according to USP/EP

Control strain	Medium inoculation level	Incubation Conditions	Recovery Specifications	Recovery Results	Cultural Response
<i>C. albicans</i> ATCC 10231	10-100 viable microorganisms	48-120 h at 22.5 ± 2.5°C	70%≤R%≤200%	<b>84,7%</b>	Conforms
<i>A. brasiliensis</i> ATCC 16404	10-100 viable microorganisms	48-120 h at 22.5 ± 2.5°C	70%≤R%≤200%	<b>110,4%</b>	Conforms

Start analysis	07 Mar 2024	End analysis	22 Mar 2024
Quality Control	Elisabetta Peri	<i>Elisabetta Peri</i>	
Quality Assurance	Alessia Diego	<i>Alessia Diego</i>	

R55480102

 11-Mar-24 8:37:34 GMT  
 Page 1 of 1

<b>Customer Name:</b>	CPC Biotech S.r.L.	<b>Processing Facility:</b>	Minerbio	<b>Work Order #</b>	3515345
<b>P.O.#</b>	84050			<b>Sales Order #</b>	3508510
	<b>10-25 kGy</b>	<b>CPCB gamma treatment</b>	<b>Irradiation Date/Time:</b>	8-Mar-24 11:26:31 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900759	DDT 84050
			270_CPCB_870/20	Std cycle 15kGy		
102.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900760	DDT 84050
			270_CPCB_870/20	Std cycle 15kGy		
103.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900761	DDT 84050
			270_CPCB_870/20	Std cycle 15kGy		
104.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900762	DDT 84050
			270_CPCB_870/20	Std cycle 15kGy		
	<b>4</b>	<b>PL</b>	<b>Total</b>			

### Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail
450.00	Minimum Dose	10.0 kGy	25.0 kGy	15.5 KGY	Pass
450.00	Maximum Dose	10.0 kGy	25.0 kGy	19.4 KGY	Pass

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

 Electronically Signed By: VALERIA LOSITO  
 Reason: Work Order Completions

Date: 11-Mar-24 08:36:18 GMT