

# Certificate of Analysis



Lot status: **APPROVED**

<b>MODEL</b> Petri ø90mm Triple Wrapped Sterile Irradiated	<b>PRODUCT CODE</b>	<b>449GLAC1000/22</b>
<b>MEDIA</b> TSA+Lec+T80+Gly+LCT1000	<b>MEDIA COMPOSITION</b>	<b>Tryptone Soy Agar + 0.07% Lecithin + 0.5% Tween80 + 0.2% Glycine + 0.3% Pyruvate + 1000U.I./Lt Lactamator TM</b>
<b>BATCH</b> 97401	<b>STORAGE CONDITIONS</b>	<b>at 2-25°C until expiry date</b>
<b>DATE OF MANUFACTURING</b> 12 Jan 2024	<b>EXPIRY DATE</b>	<b>12 Sep 2024</b>

Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	light amber coloured, clear to slightly opalescent gel forms in plates	Conforms
pH specification	7.3±0.2	Conforms
Filling media volume	30ml±1ml	Conforms
Packaging	Triple Wrapped Sterile Irradiated (T.W.S.I.)	Conforms
Dose of irradiation	10-25 KGy	Conforms Irradiation certificate Nr: 3486048

## MICROBIOLOGICAL TEST CONTROLS

- Sterility control**

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 120 hours	No growth	Conforms
22.5±2.5°C aerobic for 120 hours		

- Fertility – Growth Promotion Test** according to USP/EP

Control strain	Medium inoculation level	Incubation Conditions	Recovery Specifications	Recovery Results	Cultural Response
<i>E. coli</i> ATCC 8739	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	100,0%	Conforms
<i>P. aeruginosa</i> ATCC 9027	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	97,7%	Conforms
<i>S. aureus</i> ATCC 6538	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	86,5%	Conforms
<i>B. subtilis</i> ATCC 6633	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	72,7%	Conforms
<i>C. albicans</i> ATCC 10231	10-100 viable microorganisms	72-120 h at 32.5 ± 2.5°C	70%≤R%≤200%	102,9%	Conforms
<i>A. brasiliensis</i> ATCC 16404	10-100 viable microorganisms	72-120 h at 32.5 ± 2.5°C	70%≤R%≤200%	96,3%	Conforms

- Enzymatic Activity Test**

Control strain	Incubation Conditions	Antibiotics	Response
<i>S. aureus</i> ATCC 6538	24 h at 32.5 ± 2.5°C	Penicillin 10IU	Conforms
		Meropenem 10µg	Conforms
		Cefuroxime 30µg	Conforms
		Ceftazidime 30µg	Conforms
		Cefoxitin 30µg	Conforms

<b>Start analysis</b>	<b>12 Jan 2024</b>	<b>End analysis</b>	<b>05 Feb 2024</b>
<b>Quality Control</b>	<b>Elisabetta Peri</b>	<i>Elisabetta Peri</i>	
<b>Quality Assurance</b>	<b>Celeste Annovazzi</b>	<i>Celeste Annovazzi</i>	

R55480102

 22-Jan-24 9:58:53 GMT  
 Page 1 of 1

<b>Customer Name:</b>	CPC Biotech S.r.L.	<b>Processing Facility:</b>	Minerbio	<b>Work Order #</b>	3486048
<b>P.O.#</b>	84016			<b>Sales Order #</b>	3474930
	<b>10-25 kGy</b>	<b>CPCB gamma treatment</b>	<b>Irradiation Date/Time:</b>		19-Jan-24 12:12:39 GMT

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900729	DDT 84016
			270_CPCB_870/20	Std cycle 15kGy		
102.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900730	DDT 84016
			270_CPCB_870/20	Std cycle 15kGy		
103.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900731	DDT 84016
			270_CPCB_870/20	Std cycle 15kGy		
104.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900732	DDT 84016
			270_CPCB_870/20	Std cycle 15kGy		
105.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900733	DDT 84016
			270_CPCB_870/20	Std cycle 15kGy		
106.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900734	DDT 84016
			270_CPCB_870/20	Std cycle 15kGy		
	<b>6</b>	<b>PL</b>	<b>Total</b>			

### Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail
450.00	Minimum Dose	10.0 kGy	25.0 kGy	15.3 KGY	Pass
450.00	Maximum Dose	10.0 kGy	25.0 kGy	20.8 KGY	Pass

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

 Electronically Signed By: MILENA TAMPELLINI  
 Reason: Work Order Completions

Date: 22-Jan-24 08:13:59 GMT