

Certificate of Analysis



Lot status: **APPROVED**

MODEL Petri ø90mm Triple Wrapped Sterile Irradiated	PRODUCT CODE 449INHY/22
MEDIA TSA+Lec+T80+Hys+ThioS	MEDIA COMPOSITION Tryptone Soy Agar + 0.07% Lecithin + 0.5% Tween80 + 0.1% L-Hystidine + 0.05% S-Thiosulphate
BATCH 87401	STORAGE CONDITIONS at 2-25°C until expiry date
DATE OF MANUFACTURING 16 Jan 2024	EXPIRY DATE 16 Sep 2024

Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	light amber coloured, clear to slightly opalescent gel forms in plates	Conforms
pH specification	7.3±0.2	Conforms
Filling media volume	30ml±1ml	Conforms
Packaging	Triple Wrapped Sterile Irradiated (T.W.S.I.)	Conforms
Dose of irradiation	10-25 KGy	Conforms Irradiation certificate Nr: 3486048

MICROBIOLOGICAL TEST CONTROLS

- Sterility control**

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 120 hours 22.5±2.5°C aerobic for 120 hours	No growth	Conforms

- Fertility – Growth Promotion Test** according to USP/EP

Control strain	Medium inoculation level	Incubation Conditions	Recovery Specifications	Recovery Results	Cultural Response
<i>E. coli</i> ATCC 8739	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	105,8%	Conforms
<i>P. aeruginosa</i> ATCC 9027	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	93,2%	Conforms
<i>S. aureus</i> ATCC 6538	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	78,4%	Conforms
<i>B. subtilis</i> ATCC 6633	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	97,3%	Conforms
<i>C. albicans</i> ATCC 10231	10-100 viable microorganisms	72-120 h at 32.5 ± 2.5°C	70%≤R%≤200%	91,9%	Conforms
<i>A. brasiliensis</i> ATCC 16404	10-100 viable microorganisms	72-120 h at 32.5 ± 2.5°C	70%≤R%≤200%	95,7%	Conforms

Start analysis	16 Jan 2024	End analysis	05 Feb 2024
Quality Control	Elisabetta Peri	<i>Elisabetta Peri</i>	
Quality Assurance	Celeste Annovazzi	<i>Celeste Annovazzi</i>	

R55480102

22-Jan-24 9:58:53 GMT
Page 1 of 1

Customer Name:	CPC Biotech S.r.L.	Processing Facility:	Minerbio	Work Order #	3486048
P.O.#	84016			Sales Order #	3474930
	10-25 kGy	CPCB gamma treatment	Irradiation Date/Time:	19-Jan-24 12:12:39 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900729	DDT 84016
		Dose Map	270_CPCB_870/20	Std cycle 15kGy		
102.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900730	DDT 84016
		Dose Map	270_CPCB_870/20	Std cycle 15kGy		
103.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900731	DDT 84016
		Dose Map	270_CPCB_870/20	Std cycle 15kGy		
104.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900732	DDT 84016
		Dose Map	270_CPCB_870/20	Std cycle 15kGy		
105.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900733	DDT 84016
		Dose Map	270_CPCB_870/20	Std cycle 15kGy		
106.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900734	DDT 84016
		Dose Map	270_CPCB_870/20	Std cycle 15kGy		
	6	PL	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail
450.00	Minimum Dose	10.0 kGy	25.0 kGy	15.3 KGY	Pass
450.00	Maximum Dose	10.0 kGy	25.0 kGy	20.8 KGY	Pass

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: MILENA TAMPELLINI
Reason: Work Order Completions

Date: 22-Jan-24 08:13:59 GMT