

# Certificate of Analysis



Lot status: APPROVED

<b>MODEL</b> RODAC ø55mm Triple Wrapped Sterile Irradiated	<b>PRODUCT CODE</b>	449LAC1000/21
<b>MEDIA</b> TSA+Lec+T80+LCT1000	<b>MEDIA COMPOSITION</b>	Tryptone Soy Agar + 0.07% Lecithin + 0.5% Tween80 + 1000U.I./Lt Lactamator TM
<b>BATCH</b> 38309	<b>STORAGE CONDITIONS</b>	at 2-25°C until expiry date
<b>DATE OF MANUFACTURING</b> 13 Nov 2023	<b>EXPIRY DATE</b>	13 Jul 2024

Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	light amber coloured, clear to slightly opalescent gel forms in plates	Conforms
pH specification	7.3±0.2	Conforms
Filling media volume	17ml±1ml	Conforms
Packaging	Triple Wrapped Sterile Irradiated (T.W.S.I.)	Conforms
Dose of irradiation	10-25 KGy	Conforms Irradiation certificate Nr:3453777

## MICROBIOLOGICAL TEST CONTROLS

- Sterility control**

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 120 hours 22.5±2.5°C aerobic for 120 hours	No growth	Conforms

- Fertility – Growth Promotion Test** according to USP/EP

Control strain	Medium inoculation level	Incubation Conditions	Recovery Specifications	Recovery Results	Cultural Response
<i>E. coli</i> ATCC 8739	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	97,3%	Conforms
<i>P. aeruginosa</i> ATCC 9027	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	76,2%	Conforms
<i>S. aureus</i> ATCC 6538	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	82,5%	Conforms
<i>B. subtilis</i> ATCC 6633	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	94,9%	Conforms
<i>C. albicans</i> ATCC 10231	10-100 viable microorganisms	72-120 h at 32.5 ± 2.5°C	70%≤R%≤200%	112,8%	Conforms
<i>A. brasiliensis</i> ATCC 16404	10-100 viable microorganisms	72-120 h at 32.5 ± 2.5°C	70%≤R%≤200%	84,0%	Conforms

- Enzimatic Activity Test**

Control strain	Incubation Conditions	Antibiotics	Response
<i>S. aureus</i> ATCC 6538	24 h at 32.5 ± 2.5°C	Penicillin 10IU	Conforms
		Meropenem 10µg	Conforms
		Cefuroxime 30µg	Conforms
		Ceftazidime 30µg	Conforms
		Cefoxitin 30µg	Conforms

Start analysis	14 Nov 2023	End analysis	04 Dec 2023
Quality Control	Elisabetta Peri	<i>Elisabetta Peri</i>	
Quality Assurance	Celeste Annovazzi	<i>Celeste Annovazzi</i>	