

Certificate of Analysis



Lot status: APPROVED

MODEL RODAC ø55mm Triple Wrapped Sterile Irradiated	PRODUCT CODE	449LAC1000/21
MEDIA TSA+Lec+T80+LCT1000	MEDIA COMPOSITION	Tryptone Soy Agar + 0.07% Lecithin + 0.5% Tween80 + 1000U.I./Lt Lactamator TM
BATCH 38310	STORAGE CONDITIONS	at 2-25°C until expiry date
DATE OF MANUFACTURING 19 Dec 2023	EXPIRY DATE	19 Aug 2024

Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	light amber coloured, clear to slightly opalescent gel forms in plates	Conforms
pH specification	7.3±0.2	Conforms
Filling media volume	17ml±1ml	Conforms
Packaging	Triple Wrapped Sterile Irradiated (T.W.S.I.)	Conforms
Dose of irradiation	10-25 KGy	Conforms Irradiation certificate Nr: 3482280

MICROBIOLOGICAL TEST CONTROLS

- Sterility control**

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 120 hours	No growth	Conforms
22.5±2.5°C aerobic for 120 hours		

- Fertility – Growth Promotion Test** according to USP/EP

Control strain	Medium inoculation level	Incubation Conditions	Recovery Specifications	Recovery Results	Cultural Response
<i>E. coli</i> ATCC 8739	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	82,5%	Conforms
<i>P. aeruginosa</i> ATCC 9027	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	80,0%	Conforms
<i>S. aureus</i> ATCC 6538	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	81,1%	Conforms
<i>B. subtilis</i> ATCC 6633	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	70%≤R%≤200%	89,5%	Conforms
<i>C. albicans</i> ATCC 10231	10-100 viable microorganisms	72-120 h at 32.5 ± 2.5°C	70%≤R%≤200%	74,5%	Conforms
<i>A. brasiliensis</i> ATCC 16404	10-100 viable microorganisms	72-120 h at 32.5 ± 2.5°C	70%≤R%≤200%	73,0%	Conforms

- Enzymatic Activity Test**

Control strain	Incubation Conditions	Antibiotics	Response
<i>S. aureus</i> ATCC 6538	24 h at 32.5 ± 2.5°C	Penicillin 10IU	Conforms
		Meropenem 10µg	Conforms
		Cefuroxime 30µg	Conforms
		Ceftazidime 30µg	Conforms
		Cefoxitin 30µg	Conforms

Start analysis	19 Dec 2023	End analysis	29 Jan 2024
Quality Control	Elisabetta Peri		
Quality Assurance	Celeste Annovazzi		

R55480102

4-Jan-24 6:55:39 GMT
Page 1 of 1

Customer Name:	CPC Biotech S.r.L.	Processing Facility:	Minerbio	Work Order #	3474134
P.O.#	80224			Sales Order #	3460338
	10-25 kGy	CPCB gamma treatment	Irradiation Date/Time:	22-Dec-23 13:16:24 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900717	DDT 80224
		Dose Map	270_CPCB_870/20	Std cycle 15kGy		
102.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900718	DDT 80224
		Dose Map	270_CPCB_870/20	Std cycle 15kGy		
103.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900719	DDT 80224
		Dose Map	270_CPCB_870/20	Std cycle 15kGy		
104.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900720	DDT 80224
		Dose Map	270_CPCB_870/20	Std cycle 15kGy		
105.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900721	DDT 80224
		Dose Map	270_CPCB_870/20	Std cycle 15kGy		
106.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900722	DDT 80224
		Dose Map	270_CPCB_870/20	Std cycle 15kGy		
	6	PL	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail
450.00	Minimum Dose	10.0 kGy	25.0 kGy	15.6 KGY	Pass
450.00	Maximum Dose	10.0 kGy	25.0 kGy	20.4 KGY	Pass

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

Electronically Signed By: VALERIA LOSITO
Reason: Work Order Completions

Date: 4-Jan-24 06:54:02 GMT

R55480102

 15-Jan-24 9:18:00 GMT
 Page 1 of 1

Customer Name:	CPC Biotech S.r.L.	Processing Facility:	Minerbio	Work Order #	3482280
P.O.#	84003			Sales Order #	3470691
	10-25 kGy	CPCB gamma treatment	Irradiation Date/Time:	12-Jan-24 11:32:01 GMT	

SO Line #	Qty	UOM	Customer Item Number	Customer Item Description	Customer Lot Number	Customer Load Number
101.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900725	DDT 84003
			270_CPCB_870/20	Std cycle 15kGy		
102.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900726	DDT 84003
			270_CPCB_870/20	Std cycle 15kGy		
103.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900727	DDT 84003
			270_CPCB_870/20	Std cycle 15kGy		
104.000	1	PL	2684/20_870/20	Microbiological culture media	WWW0539900728	DDT 84003
			270_CPCB_870/20	Std cycle 15kGy		
	4	PL	Total			

Quality Test Summary

Op#	Quality Test Description	Minimum Spec	Maximum Spec	Result	Pass/Fail
450.00	Minimum Dose	10.0 kGy	25.0 kGy	15.3 KGY	Pass
450.00	Maximum Dose	10.0 kGy	25.0 kGy	19.4 KGY	Pass

Sterigenics certifies that the materials listed above (as described by the Manufacturer) received the indicated doses within the precision and accuracy of the dosimetry system employed.

 Electronically Signed By: VALERIA LOSITO
 Reason: Work Order Completions

Date: 15-Jan-24 09:16:06 GMT