Certificate of Analysis



Lot status: APPROVED

MODEL: MEDIA:	100ml w/Pierceable Flip-Off Cap Bottle Fluid Thioglycollate Medium	PRODUCT CODE:
BATCH:	23F2202	STORAGE CONDITIONS:
DATE OF MA	NUFACTURING: 22 Jun 2023	EXPIRY DATE:

ME009/30FLP32.100

at 2-25°C until expiry date

22 Jun 2024

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	yellow straw coloured, clear to very slightly opalescent solution with upper 10% or less medium pink on standing.	Conforms
pH specification	7.1±0.2	Conforms
Filling media volume	100ml	Conforms
Media composition	Fluid Thioglycollate Medium Medium is prepared utilizing raw materials declared to be TSE-BSE FREE by the manufacturer	Conforms

MICROBIOLOGICAL TEST CONTROLS

<u>Sterility control</u>

TESTS	SPECIFICATION	RESULT	
32.5±2.5°C aerobic for 336 hours	No growth	Conforme	
22.5±2.5°C aerobic for 336 hours	No growth	Conforms	

• <u>Fertility – Growth Promotion Test</u> according to USP/EP

Control strain	Medium inoculation level	Incubation Conditions	Recovery Specifications	Recovery Results	Cultural Response
P. aeruginosa ATCC 9027	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	Positive growth	Positive growth	Conforms
C. sporogenes ATCC 11437	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	Positive growth	Positive growth	Conforms
<i>S. aureus</i> ATCC 6538	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	Positive growth	Positive growth	Conforms

Start analisys	23 Jun 2023	End analisys	07 Jul 2023
Quality Control	Claudia Previtali	Claudia Reviter	
Quality Assurance	Rosaria Cassese	lonea	2 Conere

CPC Biotech s.r.l. Via per Ossona, 22 20010 Casorezzo (MI) Italy	Tel. +39 039 6898029 Fax. +39 039 6899774 www.cpcbiotech.it	CPC_ME009/30FLP32.100	Page 1/1
--	---	-----------------------	----------