Certificate of Analysis



Lot status: APPROVED

MODEL: MEDIA:	100ml Bottle with Pierceable Cap Fluid Thioglycollate Medium	PRODUCT CODE:	ME009/30P.100
BATCH:	23H0102	STORAGE CONDITIONS:	at 2-25°C until expiry date
DATE OF MA	NUFACTURING: 01 Aug 2023	EXPIRY DATE:	01 Aug 2024

CHEMICAL/PHYSICAL TESTS	SPECIFICATION	RESULT
Appearance	yellow straw coloured, clear to very slightly opalescent solution with upper 10% or less medium pink on standing.	Conforms
pH specification	7.1±0.2	Conforms
Filling media volume	100ml	Conforms
	Fluid Thioglycollate Medium	
Media composition	Medium is prepared utilizing raw materials	Conforms
	declared to be TSE-BSE FREE by the manufacturer	

MICROBIOLOGICAL TEST CONTROLS

<u>Sterility control</u>

TESTS	SPECIFICATION	RESULT
32.5±2.5°C aerobic for 336 hours	No growth	Conforms
22.5±2.5°C aerobic for 336 hours	ivo giowui	Conjorms

• <u>Fertility – Growth Promotion Test</u> according to USP/EP

Control strain	Medium inoculation level	Incubation Conditions	Recovery Specifications	Recovery Results	Cultural Response
P. aeruginosa ATCC 9027	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	Positive growth	Positive growth	Conforms
C. sporogenes ATCC 11437	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	Positive growth	Positive growth	Conforms
<i>S. aureus</i> ATCC 6538	10-100 viable microorganisms	24-72 h at 32.5 ± 2.5°C	Positive growth	Positive growth	Conforms

Start analisys	02 Aug 2023	End analisys	21 Aug 2023
Quality Control	Claudia Previtali	Claudiarreviter	
Quality Assurance	Rosaria Cassese	lonex	2 Contre

GHIARONI CODE: 953808006

CPC Biotech s.r.l.	Tel. +39 039 6898029		
Via per Ossona, 22	Fax. +39 039 6899774	CPC_ME009/30P.100	Page 1/1
20010 Casorezzo (MI) Italy	www.cpcbiotech.it		8